


# METROLOGIA

J3/2003

60 MPa, average pressure



***Intercomparison in gauge pressure  
range 0 ... 60 MPa***

Markku Rantanen

Helsinki 2003



**MITTATEKNIIKAN KESKUS**  
**CENTRE FOR METROLOGY AND ACCREDITATION**

**Julkaisu J3/2003**

**INTERCOMPARISON IN GAUGE PRESSURE RANGE 0 ... 60 MPa**

**Markku Rantanen**

**Helsinki 2003**

## **ABSTRACT**

An intercomparison P14 in the gauge pressure range 0 ... 60 MPa was arranged in 2003 by the Centre for Metrology and Accreditation (MIKES). Five pressure calibration laboratories participated in the comparison, four laboratories from Finland and one from Sweden).

The transfer standard was a Beamex MC5 pressure calibrator equipped with an external 60 MPa sensor. The participants were advised to use their routine methods and existing procedures in calibrating the transfer standard. So it was possible to compare not only pressure results but also the measurement methods and the contents of the calibration certificates.

All the pressure results were in a good agreement with the results of MIKES.

## **TIIVISTELMÄ**

Mittateknikan keskus (MIKES) järjesti vuonna 2003 vertailumittauksen P14 ylipainealueella 0 ... 60 MPa. Vertailumittaukseen osallistui neljä kalibrointilaboratorioa Suomesta ja yksi Ruotsista.

Kiertävä vertailulaite oli Beamex MC5 painekalibraattori varustettuna ulkoisella anturilla. Mittausohjeissa kehotettiin osanottajia tekemään mittaukset omien rutiinimenetelyjensä mukaisesti. Näin voitiin varsinaisten paineen mittaustulosten lisäksi vertailla mittausmenetelmiä ja kalibrointitodistusten sisältöä.

Kaikki paineen mittaustulokset olivat mittausepävarmuksien puitteissa samoja kuin MIKESin tulokset.

## CONTENTS

### ABSTRACT TIIVISTELMÄ

1. INTRODUCTION	7
2. TRANSFER STANDARD	7
3. REFERENCE LABORATORY	7
4. PARTICIPANTS	8
5. REFERENCE VALUES	8
6. MEASUREMENT INSTRUCTIONS	8
7. RESULTS	9
8. MEASUREMENT PROCEDURES AND CONTENTS OF CERTIFICATES	10
9. CONCLUSIONS	10
10. REFERENCES	10
Figures 1 to 6 on results	11
Table 1: Comparison of calibration methods and certificate contents	14
Appendix 1: Summary of results from each participating laboratory	15

## **INTERCOMPARISON IN GAUGE PRESSURE RANGE 0 ... 60 MPa**

### **1 INTRODUCTION**

An intercomparison P14 was arranged in 2003 by the Centre for Metrology and Accreditation (MIKES) in the gauge pressure range 0 ... 60 MPa.

The aim of the comparison was to compare not only the results of pressure measurements but also the methods in making the measurements and the expression of the results and uncertainties in a routine calibration job. The transfer standard used in the comparison was an instrument familiar to all of the participants. No detailed instructions were given to the laboratories but they were advised to use their routine methods.

Four accredited pressure calibration laboratories from Finland and one from Sweden participated in the comparison.

### **2 TRANSFER STANDARD**

The transfer standard was a Beamex MC5 pressure calibrator s/n 24512824 equipped with an external sensor Beamex EXT 600 s/n 20854, made available by Oy Beamex Ab. According to the manufacturer, the sensor is intended to be used in the horizontal position. Any non-corrosive gas or liquid can be used. For the instrument as a standard product the maximum error is defined as 0,04% of reading + 0,015% of full scale.

Before the circulation, the Beamex laboratory adjusted a small pressure dependent error in the transfer standard. However, the output was within the specifications.

### **3 REFERENCE LABORATORY**

For the oil pressure range 0,5 MPa ... 500 MPa the reference standard of MIKES is a Desgranges & Huot 5306 pressure balance. This pressure range is covered by four piston-cylinder units. The effective areas are traceable to Laboratoire National d'Essais (BNM-LNE), Paris. The masses of the ring weights were determined by the mass laboratory of MIKES.

**4****PARTICIPANTS**

The following laboratories participated in the comparison:

Finnair Oy Test Equipment Calibration, Vantaa  
 Finnish Air Force, Air Materiel Command, Calibration Laboratory, Tampere  
 Inspecta Oy Measuring Technology, Helsinki  
 Oy Beamex Ab, Calibration Laboratory, Pietarsaari  
 SAAB AB, Linköping, Sweden

**5****CALCULATION OF THE REFERENCE VALUES**

The transfer standard was calibrated three times at the reference laboratory MIKES. The first calibration was made at beginning of May and the last at the end of June in 2003.

Figure 1 shows all the result points of MIKES calibrations. The stability of the transfer standard was very good, and drift during the circulation was negligible.

The reference values were calculated as the averages of the MIKES results for each nominal pressure in two ways. The first way was to calculate the reference values separately for increasing and decreasing pressures, and the second to calculate only one reference value for each nominal pressure. In the latter case the uncertainties are higher due to the hysteresis effects.

The major uncertainty components taken into account were

- uncertainty in setting the zero of the transfer standard
- resolution of the transfer standard
- scatter of the results
- hysteresis (not for the separate values for increasing and decreasing pressures)
- uncertainty of the measurement standard.

According to the manufacturer the effects of calibration temperature can be neglected in a typical laboratory environment.

**6****MEASUREMENT INSTRUCTIONS**

The measurement instructions were short and straightforward: The participants were asked to calibrate the transfer standard using their own procedures. However, the participants were advised to select MPa for the pressure unit, and the use of the nominal pressures 5 MPa, 10 MPa, 20 MPa, 30 MPa, 40 MPa, 50 MPa and 60 MPa was recommended.

Finally, the participants were asked to send their results as calibration certificates to MIKES within two weeks after the measurements.

## RESULTS

All the laboratories presented their results as calibration certificates, and indicated that the uncertainty of calibration was estimated using the document EA-4/02 and a coverage factor  $k = 2$ .

Following the EA intercomparison practice all laboratories were given letter codes. Each laboratory knows only its own code.

A summary of all the result is shown in Appendix 1. The figures 2 to 6 show results at nominal pressures 5 MPa up, 5 MPa down, 30 MPa up, 30 MPa average and 60 MPa.

A tool often used in analysing results from interlaboratory comparisons is the normalised error  $E_n$ , which takes into account both the result and its uncertainty. The normalised error  $E_n$  is calculated as

$$E_n = \frac{(p_{transfer} - p_{std})_{lab} - (p_{transfer} - p_{std})_{ref}}{\sqrt{(U_{lab})^2 + (U_{ref})^2}}$$

where  $p_{transfer}$  is pressure indicated by the transfer standard,  
 $p_{std}$  is the pressure of the laboratory standard,  
 $U_{lab}$  is the uncertainty of the laboratory result, and  
 $U_{ref}$  is the uncertainty of the reference value.

The  $E_n$ -values of all the results are shown in Appendix 1.

A summary of the  $E_n$ -values is in the following:

Laboratory	range of $E_n$ -values
<b>A</b>	-0,34 ... 0,80
<b>B</b>	-0,77 ... -0,16
<b>C</b>	-0,27 ... 0,21
<b>D</b>	-0,01 ... 0,43
<b>E</b>	0,02 ... 0,53

The result in an interlaboratory comparison is regarded as correct within the limits of uncertainty, if the absolute value of the normalised error  $E_n$  is less than 1.

In this case the  $E_n$ -values for all the results from all the laboratories are between -1 and +1.

The absolute  $E_n$ -values above 0,5 seem to be more frequent at lower nominal pressures. The reason may be the underestimation or neglecting the uncertainty in setting the transfer standard zero.

## MEASUREMENT PROCEDURES AND CONTENTS OF CERTIFICATES

As the measurement procedure was not specified, there was variation in the methods used by the participants. A comparison of the calibration methods and the certificate contents is shown in Table 1.

The question marks in the Table 1 mean that the information is not directly shown in the certificate. References to internal laboratory procedures may contain more information but it does not immediately help the reader.

Generally speaking the certificates of all the participating laboratories were in accordance with the requirements of ISO/IEC 17025. However, a remark can be made on the ISO/IEC 17025 point 5.10.4.1 a. This point states that calibration certificates should include *the conditions under which the calibrations were made that have influence on the measurement results*. The position of the pressure sensor, and the procedures applied in pre-pressurising and in setting the zero are potential influence factors in pressure calibrations. Including this information in the certificate is highly recommended.

## CONCLUSIONS

Five pressure calibration laboratories participated in an intercomparison in the gauge pressure range 0 to 60 MPa between April and June 2003.

All the results from all the participating laboratories were in a good agreement with the reference values from MIKES. However, there was some variation in the measurement methods and in the contents of the calibration certificates.

The transfer standard, a Beamex MC5 calibrator with an external pressure sensor EXT600, was found to be very stable and its resolution was good enough for an intercomparison at this level.

## REFERENCES

EA-4/02: Expression of the Uncertainty of Measurement in Calibration



ISO/IEC 17025: General requirements for the testing and calibration laboratories.



Certificates of Calibration M-03P060, M-03P069 and M-03P084. Centre for Metrology and Accreditation



MR 25.9.2003

**Fig. 1. Beamex EXT 600 stability**

Calibrations at MIKES (all results)  
Deviations from the straight line fitted to M1 results.

**Fig. 2. 5 MPa, increasing pressure**

**Fig. 3. 5 MPa, decreasing pressure****Fig. 4. 30 MPa, increasing pressure**

**Fig. 5. 30 MPa, average pressure****Fig. 6. 60 MPa, average pressure**

**Table 1. Comparison of calibration methods and certificate contents**

Laboratory code	A	B	C	D	E	M2	M3
<b>Ambient temperature</b>	21+-0.5	23+-1	23+-2	23+-2	21.5	22.4+-1	23+-1
<b>Sensor position</b>	Vertical, connection down	?	?	?	Vertical, connection up	Horizontal	Horizontal
<b>Pressure medium</b>	Oil	Oil	Distilled water	Oil	Oil	Oil	Oil
<b>Pre-pressures</b>	?	60 MPa twice	60 MPa 3 times	?	60 MPa once	60 MPa twice	60 MPa twice
<b>Zero setting</b>	Before each cycle	?	?	?	After pre- pressures	After pre- pressures	After pre- pressures
<b>Unit of pressure</b>	MPa	MPa	MPa	MPa	MPa	MPa	MPa
<b>N:o of nominal pressures up down</b>	11 10	8 7	8 7	8 7	8 7	8 7	8 7
<b>N:o of measurements at each level</b>							
up	3	3	2	4	1	2	3
down	3	3	2	4	1	2	3
<b>Total n:o of measurements</b>	63	45	30	60	15	30	45
<b>Result as error or correction</b>	error	error	error	averages of ref. and obj.readings	error	error	error
<b>Results up/down</b>	average	yes	yes	yes	yes	average	yes
<b>Results as graph</b>	no	no	yes	no	no	yes	yes
<b>Uncertainty components named</b>	yes	yes	yes	no	yes	yes	yes
<b>Values for components</b>	no	no	yes	no	no	no	no
<b>Uncertainty at 60 MPa (k=2), MPa parts per million</b>	0.009 150	0.0079 132	0.0125 208	0.03 500	0.018 300	0.003 50	0.003 50

**APPENDIX 1.**  
**Summary of results in pressure intercomparison P14**

**Page 1(3)**

Lab code Nominal pressure MPa	A		Ref. value MPa	Ref. uncertainty MPa	Deviation from ref. MPa	E(n)
Lab code Nominal pressure MPa	Result MPa	Uncert. MPa	Ref. value MPa	Ref. uncertainty MPa	Deviation from ref. MPa	E(n)
5	0.003	0.0014	0.0013	0.0016	0.0017	0.80
10	0.005	0.0017	0.0033	0.0017	0.0017	0.71
20	0.010	0.0030	0.0084	0.0019	0.0016	0.45
30	0.014	0.0045	0.0134	0.0022	0.0006	0.12
40	0.018	0.0060	0.0190	0.0026	-0.0010	-0.15
50	0.022	0.0075	0.0245	0.0029	-0.0025	-0.31
60	0.026	0.0090	0.0292	0.0030	-0.0032	-0.34
Lab code Nominal pressure MPa	B		Ref. value MPa	Ref. uncertainty MPa	Deviation from ref. MPa	E(n)
5 up	0.0004	0.0018	0.002	0.0012	-0.0016	-0.74
10 up	0.0026	0.0022	0.0043	0.0012	-0.0017	-0.68
20 up	0.0068	0.0029	0.0093	0.0015	-0.0025	-0.77
30 up	0.0116	0.0041	0.0144	0.0017	-0.0028	-0.63
40 up	0.0170	0.0054	0.0202	0.0021	-0.0032	-0.55
50 up	0.0222	0.0067	0.0256	0.0025	-0.0034	-0.48
60 up	0.0267	0.0079	0.0296	0.0028	-0.0029	-0.35
50 down	0.0215	0.0067	0.0234	0.0025	-0.0019	-0.27
40 down	0.0164	0.0054	0.0178	0.0021	-0.0014	-0.24
30 down	0.0107	0.0041	0.0124	0.0017	-0.0017	-0.38
20 down	0.0061	0.0029	0.0076	0.0015	-0.0015	-0.46
10 down	0.0020	0.0022	0.0024	0.0012	-0.0004	-0.16
5 down	0.0001	0.0018	0.0006	0.0012	-0.0005	-0.23
Lab code Nominal pressure MPa	C		Ref. value MPa	Ref. uncertainty MPa	Deviation from ref. MPa	E(n)
5 up	0.0025	0.0021	0.002	0.0012	0.0005	0.21
10 up	0.0042	0.0028	0.0043	0.0012	-0.0001	-0.03
20 up	0.0093	0.0046	0.0093	0.0015	0.0000	0.00
30 up	0.0128	0.0066	0.0144	0.0017	-0.0016	-0.23
40 up	0.0188	0.0085	0.0202	0.0021	-0.0014	-0.16
50 up	0.0227	0.0105	0.0256	0.0025	-0.0029	-0.27
60 up	0.0267	0.0125	0.0296	0.0028	-0.0029	-0.23
50 down	0.0217	0.0105	0.0234	0.0025	-0.0017	-0.16
40 down	0.0168	0.0085	0.0178	0.0021	-0.0010	-0.11
30 down	0.0109	0.0066	0.0124	0.0017	-0.0015	-0.22
20 down	0.0063	0.0046	0.0076	0.0015	-0.0013	-0.27
10 down	0.0017	0.0028	0.0024	0.0012	-0.0007	-0.23
5 down	0.0010	0.0021	0.0006	0.0012	0.0004	0.17

**APPENDIX 1. continued:**  
**Summary of results in pressure intercomparison P14**

**Page 2(3)**

Lab code Nominal pressure	D		Ref. value	Ref. uncertainty	Deviation from ref.	E(n)
MPa	Result MPa	Uncert. MPa	MPa	MPa	MPa	
5 up	0.003	0.003	0.0020	0.0012	0.001	0.31
10 up	0.005	0.005	0.0043	0.0012	0.0007	0.14
20 up	0.010	0.010	0.0093	0.0015	0.0007	0.07
30 up	0.015	0.015	0.0144	0.0017	0.0006	0.04
40 up	0.020	0.020	0.0202	0.0021	-0.0002	-0.01
50 up	0.026	0.025	0.0256	0.0025	0.0004	0.02
60 up	0.032	0.030	0.0296	0.0028	0.0024	0.08
50 down	0.024	0.025	0.0234	0.0025	0.0006	0.02
40 down	0.018	0.020	0.0178	0.0021	0.0002	0.01
30 down	0.013	0.050	0.0124	0.0017	0.0006	0.01
20 down	0.008	0.010	0.0076	0.0015	0.0004	0.04
10 down	0.003	0.005	0.0024	0.0012	0.0006	0.12
5 down	0.002	0.003	0.0006	0.0012	0.0014	0.43
Lab code Nominal pressure	E		Ref. value	Ref. uncertainty	Deviation from ref.	E(n)
MPa	Result MPa	Uncert. MPa	MPa	MPa	MPa	
5 up	0.003	0.002	0.0020	0.0012	0.0010	0.43
10 up	0.006	0.003	0.0043	0.0012	0.0017	0.53
20 up	0.011	0.006	0.0093	0.0015	0.0017	0.27
30 up	0.016	0.009	0.0144	0.0017	0.0016	0.17
40 up	0.021	0.012	0.0202	0.0021	0.0008	0.07
50 up	0.026	0.015	0.0256	0.0025	0.0004	0.03
60 up	0.030	0.018	0.0296	0.0028	0.0004	0.02
50 down	0.024	0.015	0.0234	0.0025	0.0006	0.04
40 down	0.019	0.012	0.0178	0.0021	0.0012	0.10
30 down	0.014	0.009	0.0124	0.0017	0.0016	0.17
20 down	0.008	0.006	0.0076	0.0015	0.0004	0.06
10 down	0.003	0.003	0.0024	0.0012	0.0006	0.19
5 down	0.001	0.002	0.0006	0.0012	0.0004	0.17

**APPENDIX 1. continued:**  
**Summary of results in pressure intercomparison P14**

**Page 3(3)**

Lab code		MIKES 1				
Nominal pressure	Result	Uncert.	Ref. value	Ref. uncertainty	Deviation from ref.	E(n)
MPa	MPa	MPa	MPa	MPa	MPa	
5 up	0.002	0.0014	0.002	0.0012	0.0000	0.00
10 up	0.004	0.0014	0.0043	0.0012	-0.0003	-0.16
20 up	0.009	0.0016	0.0093	0.0015	-0.0003	-0.14
30 up	0.014	0.0019	0.0144	0.0017	-0.0004	-0.16
40 up	0.020	0.0022	0.0202	0.0021	-0.0002	-0.07
50 up	0.025	0.0025	0.0256	0.0025	-0.0006	-0.17
60 up	0.029	0.0029	0.0296	0.0028	-0.0006	-0.15
60 down	0.028	0.0029	0.0289	0.0029	-0.0009	-0.22
50 down	0.023	0.0025	0.0234	0.0025	-0.0004	-0.11
40 down	0.018	0.0022	0.0178	0.0021	0.0002	0.07
30 down	0.012	0.0018	0.0124	0.0017	-0.0004	-0.16
20 down	0.007	0.0016	0.0076	0.0015	-0.0006	-0.27
10 down	0.002	0.0014	0.0024	0.0012	-0.0004	-0.22
5 down	0.001	0.0014	0.0006	0.0012	0.0004	0.22
Lab code		MIKES 2				
Nominal pressure	Result	Uncert.	Ref. value	Ref. uncertainty	Deviation from ref.	E(n)
MPa	MPa	MPa	MPa	MPa	MPa	
5	0.002	0.0018	0.0013	0.0016	0.0007	0.29
10	0.004	0.0018	0.0033	0.0017	0.0007	0.28
20	0.009	0.0018	0.0084	0.0019	0.0006	0.23
30	0.014	0.0022	0.0134	0.0022	0.0006	0.19
40	0.020	0.0027	0.0190	0.0026	0.0010	0.27
50	0.026	0.0030	0.0245	0.0029	0.0015	0.36
60	0.030	0.0029	0.0292	0.0030	0.0008	0.19
Lab code		MIKES 3				
Nominal pressure	Result	Uncert.	Ref. value	Ref. uncertainty	Deviation from ref.	E(n)
MPa	MPa	MPa	MPa	MPa	MPa	
5 up	0.001	0.0014	0.002	0.0012	-0.001	-0.54
10 up	0.004	0.0014	0.0043	0.0012	-0.0003	-0.16
20 up	0.009	0.0016	0.0093	0.0015	-0.0003	-0.14
30 up	0.014	0.0019	0.0144	0.0017	-0.0004	-0.16
40 up	0.020	0.0025	0.0202	0.0021	-0.0002	-0.06
50 up	0.025	0.0025	0.0256	0.0025	-0.0006	-0.17
60 up	0.030	0.0031	0.0296	0.0028	0.0004	0.10
60 down	0.029	0.0029	0.0289	0.0029	0.0001	0.02
50 down	0.023	0.0025	0.0234	0.0025	-0.0004	-0.11
40 down	0.018	0.0022	0.0178	0.0021	0.0002	0.07
30 down	0.012	0.0019	0.0124	0.0017	-0.0004	-0.16
20 down	0.007	0.0016	0.0076	0.0015	-0.0006	-0.27
10 down	0.002	0.0014	0.0024	0.0012	-0.0004	-0.22
5 down	0.001	0.0014	0.0006	0.0012	0.0004	0.22

# Recent publications

- J5/1999 Metrologian neuvottelukunta, kemin jaosto, *Mikrobiologisen metrologian tilanneselvitys ja kehittämissuunnitelma*
- J6/1999 Finnish National Standards Laboratories FINMET. *Annual Report 1998*
- J7/1999 T. Weckström, *Lämpötilan vertailumittaus L10, S-typin termoelementin kalibointi*
- J8/1999 S. Nevalainen, *Mekaanisten väärähtelyiden mittausten kartoitus*
- J9/1999 M. Heinonen, *Intercomparison of the hydrometer calibration systems at the IMGC and the MIKES*
- J10/1999 M. Heinonen, *National basis for traceability in humidity measurements*
- J1/2000 T. Weckström, *Intercomparison of temperature standards of Lithuania and Finland*
- J2/2000 Finnish National Standards Laboratories FINMET, *Annual Report 1999*
- J3/2000 K. Riski, *Mass comparison M3*
- J4/2000 K. Riski, *Mass and volume comparisons at MIKES*
- J5/2000 A. Lassila ja S. Nevalainen, *Nanometritason mittaukset, kartoitus*
- J6/2000 M. Rantanen, *Nordic intercomparison in gauge pressure range 0 ... 2 MPa*
- J1/2001 S.I. Niemelä, *Mikrobiologian kvantitatiivisten viljelymääritysten mittausepävarmuus*
- J2/2001 J. Järvinen (Ed), *Finnish National Standards Laboratories. Annual Report 2000*
- J3/2001 T. Weckström, *Lämpötilan vertailumittaus L11, PT100-anturin sovitusmenetelmän kehittäminen*
- J4/2001 B. Hemming, *High precision roundness. Euromet project 533. Final Report*
- J5/2001 M. Heinonen, *Kaasun kosteuden mittaaaminen*
- J6/2001 M. Heinonen, S. Bell, K. Flakiewics, G. Mamontov, P.K. Birch, A. Steiner and S. Ugus, *Intercomparison of humidity standards*
- J7/2001 M. Rantanen, *Comparisons in the pressure range from 50 kPa to 350 kPa*
- J1/2002 T. Weckström, *Lämpötilan mittaus*
- J2/2002 J. Järvinen, M. Heinonen and A. Lassila (Eds), *Annual Report 2001*
- J3/2002 S.I. Niemelä, *Uncertainty of quantitative determinations derived by cultivation of microorganisms*
- J4/2002 A. Lassila, *Calibration of gauge blocks by mechanical comparison. Final Report*
- J5/2002 V. König, A. Pitkäkoski, M. Rantanen and S. Semenoja, *Comparison of spinning rotor vacuum gauges between MIKES, SP and Vaisala Oyj*
- J6/2002 M. Rantanen and S. Semenoja, *Calibration of a 130 Pa CDG: Comparison of the results from MIKES, PTB and MKS Deutschland*
- J1/2003 J. Järvinen, M. Heinonen and A. Lassila (Eds), *Annual Report 2002*
- J2/2003 K. Riski, *Basic formula for mass calibration*
- J3/2003 M. Rantanen, *Intercomparison in gauge pressure range 0...60 MPa*



- PL 239, Lönnrotinkatu 37, 00181 HELSINKI
- Puh. (09) 616761 • Fax (09) 6167467
- [www.mikes.fi](http://www.mikes.fi)